



USER GUIDE



ABDOMINAL EXAMINATION TRAINER

This product is available
in light and dark skin tones

60000

60054

FOR MORE SKILLS TRAINING PRODUCTS VISIT

limbsandthings.com

Sussex Street, St Philips, Bristol, BS2 0RA, UK | sales@limbsandthings.com | +44 (0) 117 311 0500

INDEX

PACKAGE SUPPLIED	3
SAFETY & CLEANING INSTRUCTIONS	4
TEST CERTIFICATIONS	6
SET UP	8
• ADDING/REMOVING ABDOMINAL SKIN	8
• PLACING ORGANS & PATHOLOGIES	8
• INSTALLING/CHANGING BATTERIES	12
• USING MAINS POWER	13
• SPEAKER & MP3 PLAYER	13
• INSERTING THE ABDOMINAL WALL	15
• ATTACHING THE RECTUS BAG	15
• ASCITES SET UP	16
• GASEOUS DISTENSION	17
• USING THE BREATHING MECHANISM	19
EXAMPLES OF SET UP FOR SCENARIOS	20

TAKE A LOOK
AT OUR **VIDEO**



VISIT THE WEBSITE
limbsandthings.com

OR YOUTUBE CHANNEL
youtube.com/limbsandthings

PACKAGE
SUPPLIED

Base
Unit

60005



Abdominal Examination
Skin

60006 60057



Abdominal
Rectus Bag

60007



Abdominal Examination
Organ Locator

60008



Abdominal
Wall

60009



Abdominal Examination
Livers (Set of 3)

60010




Abdominal Examination
Spleens (Set of 2)

60011



Abdominal Examination
Kidneys (x2)

60012




Abdominal Examination
Bladder

60013



Abdominal Examination
Pathologies (Set of 6)

60014



Abdominal Examination
Aortas (Set of 2)

60015



Abdominal Examination
Distension Set

60016




International 5V
DC Power Supply

60053



AA Batteries (x4)



Carry Cases
(x2)



- **Rectus Bag + Abdominal Wall:** simulates underlying layers with in the abdomen (P. 15)
- **Ascites Bag + Distension Foam:** allows for identification of distended, fluid filled abdomen by using the following techniques, percussion, shifting dullness or fluid thrill technique (P. 16)
- **Distension Bag + Distension Foam:** allows for identification of distended, air filled abdomen when percussed/palpated. A potential scenario for this would be a bowel obstruction (P. 17)



SAFETY & CLEANING INSTRUCTIONS


READ THE SAFETY INSTRUCTIONS CAREFULLY BEFORE USING YOUR ABDOMINAL EXAMINATION TRAINER

DO's



Follow all instructions.
Keep these instructions.
Heed all warnings.

Only use attachments/accessories specified by Limbs & Things.



Use only the XP Power ACM06 mains adaptor provided with the product, the use of a different adaptor can cause electrical or fire hazards.

Only used at altitudes not exceeding 5000m.



If a product is damaged or worn, it should be repaired by Limbs & Things. To arrange this, please contact your point of purchase who will be able to provide a repair quote.

Clean the product with a damp soft cloth or sponge, using only warm water with mild detergent. The product must be stored dry.



DON'Ts



Do not install near any heat sources, do not expose the product to flames or use near naked flame sources. Do not use flammable solvents near or on this product.

Never use the product outdoors as this may damage the product.



Caution - Risk of explosion if battery is replaced with incorrect type. This product is intended for use with standard type AA Alkaline or similar non-rechargeable batteries. Check local guidelines for the correct disposal of batteries.

Do not use this apparatus near water, this product contains electronics.



Do not disassemble the product or attempt to repair a faulty unit, please return the product to Limbs & Things for assessment and repair.

Do not allow young children to use this product.



SAFETY & CLEANING INSTRUCTIONS

READ THE SAFETY INSTRUCTIONS CAREFULLY BEFORE USING YOUR ABDOMINAL EXAMINATION TRAINER

STORAGE & TRANSPORTATION



The trainer and its accessories are heavy when packed in the carrying case.



Always ensure that the products are properly packed and secured during transportation and storage in order to prevent personal injury or damage to the product.

The black carry cases that the **ABDOMINAL TRAINER** is packed in should be stored flat.

✓ STORE THIS WAY



They should not be stored on their side.

✗



The cardboard box that the **ASCITES BAG** is supplied in should be kept and used to store the Ascites bag.

Always ensure both the Abdominal Rectus Bag and Abdominal Examination Ascites Bag are detached from the trainer when not in use and stored flat, without any risk of creasing/folded edges.

✓

STORE FLAT



✗



DISPOSAL INSTRUCTIONS

Do not disassemble, crush, short-circuit or incinerate the batteries as it may cause fire, injury, burns or other hazards.

Do not dispose of this product with unsorted waste. Improper disposal may be harmful to the environment and human health. Please refer to your local waste authority for information on return and collection systems in your area.



This symbol on the product(s) and / or accompanying documents means that used electrical and electronic equipment (WEEE) should not be mixed with general household waste. For proper treatment, recovery and recycling, please take this product(s) to designated collection points where it will be accepted free of charge.

Alternatively, in some countries, you may be able to return your products to your local retailer upon purchase of an equivalent new product. Disposing of this product correctly will help save valuable resources and prevent any potential negative effects on human health and the environment, which could otherwise arise from inappropriate waste handling.

Please contact your local authority for further details of your nearest designated collection point.

Penalties may be applicable for incorrect disposal of this waste, in accordance with your national legislation.

For professional users in the European Union

If you wish to discard electrical and electronic equipment (EEE), please contact your dealer or supplier for further information.

For disposal in countries outside of the European Union

This symbol is only valid in the European Union (EU). If you wish to discard this product, please contact your local authorities or dealer and ask for the correct method of disposal.



EC DECLARATION OF CONFORMITY

In accordance with ISO IEC 17050-1 : 2010

We at

Limbs & Things Ltd

of

Sussex Street, St. Philips, Bristol, BS2 0RA

Declare under our sole responsibility that the product described below:

Brand and Model Name	Limbs and Things Abdominal Examination Trainer, Part numbers: 60000, 60054, 60001, 60055.
-----------------------------	--


Has been tested and is in conformity with the following standards:

<u>SUBJECT</u>	<u>STANDARD</u>
Electrical Safety	LVD Directive – 2014/35/EU EN 60065: 2014 Eighth edition and IEC 60065: 2014 Eighth edition – Audio, Video and similar electronic apparatus – safety requirements
Electromagnetic Compatibility	EMC Directive – 2014/30/EU EN61326-1:2013 Electrical equipment for measurement, control and laboratory use. Emissions: EN55016-2-3:2006 EN61000-3-2:2006 +A1+A2 EN61000-3-3:2013 EN55011:2009+A1: 2010 Immunity: EN61000-4-2: 2009 EN61000-4-3: :2006 +A1 +A2 EN61000-4-4: 2012 EN61000-4-5: 2014 EN61000-4-6: 2014 EN61000-4-8: 2010 EN61000-4-11:2004
Radio performance	EN 300 328 V1.8.1:2012-06
RoHS	<i>RoHS - Directive 2011/65/EU, as amended (by Directive 2015/863)</i>

Thereby meeting the essential requirements of directive (RED) (2014/53/EU)

I hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced Standards and meets all essential requirements of the RTTE Directive.

Signed Nick Hull
Position Managing Director
Location Limbs & Things, Bristol, England

Signature

Date 23 August 2021



DECLARATION OF CONFORMITY

We at

Limbs & Things Ltd

of

Sussex Street, St. Philips, Bristol, BS2 0RA

Declare under our sole responsibility that the product described below:

Brand and Model Name	Limbs and Things Abdominal Examination Trainer, Part numbers: 60000, 60054, 60001, 60055.
-----------------------------	--

Has been tested and is in conformity with the following standards:

Technical regulations	FCC CFR47 Part 15
Test laboratory	Intertek Intertek House Cleeve Road Leatherhead, Surrey KT22 75B
Test report number	102856564LHD-001B
Test report date	18 th April 2017

Signed Nick Hull
Position Managing Director
Location Limbs & Things, Bristol, England


Signature 
Date 23 August 2021

SET UP

ADDING/REMOVING ABDOMINAL SKIN



Tilt the model on to its side to reveal the velcro straps







Please ensure to pull the velcro straps and NOT the skin

PLACING ORGANS & PATHOLOGIES



Remove Skin, Rectus Bag and Abdominal Wall to get to the Organ Locator






SET UP


PLACING ORGANS & PATHOLOGIES

Liver & Spleen


Select a **LIVER** from the Set of 3




Knob located on the top **RIGHT** of the model




Select a **SPLEEN** from the Set of 2



Same fitting mechanism as Liver



Knob located on the top **LEFT** of the model






SET UP



PLACING ORGANS & PATHOLOGIES

Kidney & Bladder

Remove the **KIDNEY LOCATOR**
(1 on each side of the model)



Select the **BLADDER** and fit
the 2 pegs on its back to the
holes indicated below




SET UP


PLACING ORGANS & PATHOLOGIES

Pathologies & Aortas

Various **PATHOLOGIES/MASSES** can be placed in different areas of the Organ Locator





Hard & soft variations





Each pathology/mass has 2 pegs on the bottom which can be lined up with the holes on the Organ Locator

Line up the 2 pegs on the back of the **AORTA** with the 2 holes at the highest point of the Organ Locator.




Same process for both Normal & Aneurysmal





When the batteries are low, the red light will light up

[MEDIA ERROR] - If this message displays, it means 'change the batteries'




We strongly recommend high quality Batteries (e.G. Duracell/Energizer)



Remove batteries when the model is not in use

SET UP

USING MAINS POWER



The model is also provided with a mains power lead




Use only the **XP Power ACM06** mains adaptor provided with the product, the use of a different adaptor can cause electrical or fire hazards



The unit can be used in tropical countries and **up to 5000m altitude** with the XP Power ACM06 mains adaptor


SPEAKER & MP3 PLAYER

There are 3 speaker locations on the Organ Locator




SET UP

SPEAKER & MP3 PLAYER



Rotate the **VOLUME** control to turn volume up and down



Press **POWER** button (hold for 2 seconds) and the display will light up

Press **MODE** button and ensure it is set to Normal

More sounds can be added by the user by loading them on to the provided USB, or by loading on to a soundcard (not provided)

Select the chosen sound and press **PLAY** button

Press the **ARROWS** to scroll through the different available sounds



If you add a sound file make sure it is in MP3 format and reads .mp3 at the end





If the MP3 player stops working during use, this may be due to a slight power surge. To remedy this, remove one of the batteries or unplug the USB to turn the MP3 player off, then replace and restart. If there is still no sound, please contact a member of our Customer Services Team

SET UP

INSERTING THE ABDOMINAL WALL


Once the model has been set up with the relevant organs, the **ABDOMINAL WALL** needs to be inserted



Ensure abdomen foam fits underneath liver/spleen

ATTACHING THE RECTUS BAG

Once the Abdominal Wall is in place, the **RECTUS BAG** can be fitted over the top



Secure the Rectus Bag by attaching the peg holes to the pegs on the model



STORE FLAT



Attach **SKIN** as per previous instructions (P. 8)


SET UP

ASCITES SET UP


For **ASCITES SET UP** ensure that the Organ Locator is displayed as a starting point



Ensure all organs are removed prior to insertion of distension foam




Place the **DISTENSION FOAM** on top of the Organ Locator



Use distension foam - not to be confused with the abdominal foam

Put the **ASCITES BAG** on top of the Distension Foam Insert



Secure the Ascites Bag by attaching the peg holes by the pegs on the model



STORE FLAT



Attach **SKIN** as per previous instructions (P. 8)


SET UP

GASEOUS DISTENSION


For **GASEOUS DISTENTION SET UP** ensure that the Organ Locator is displayed as a starting point




Ensure all organs are removed prior to insertion of distention foam



For **GASEOUS DISTENTION SET UP** ensure that the Organ Locator is displayed as a starting point



Place the **DISTENSION FOAM** on top of the Organ Locator.
Place the **GASEOUS DISTENSION BAG** over the Distension Foam Insert.
Secure the Gaseous Distension Bag by attaching the peg holes.




Ensure that the air inlet tube is on the underside of the Gaseous Distension Bag and under the Distension Foam. This will prevent kinks in the air inlet tube.

Attach **SKIN** as per previous instructions (P. 8)


SET UP

GASEOUS DISTENSION


Attach the blue connector from the **PUMP**



Once connected inflate the Gaseous Distension Bag with the pump.
6-8 PUMPS (in and out as 1 pump)



Disconnect the pump from the model after inflating or the Gaseous Distension Bag will deflate itself



SET UP

GASEOUS DISTENSION



EMPTYING THE GASEOUS DISTENSION BAG



USING THE BREATHING MECHANISM

When the lever is rotated to a downward position, the model simulates the patient breathing in (inhaling)


When the lever is pointed upwards, the model simulates the patient breathing out (exhaling)




The lever can be rotated 360 degrees (in either direction) to simulate continuous breathing

EXAMPLES OF
SET UP FOR
SCENARIOS


ACUTE URINARY RETENTION - BLADDER




SLIGHTLY ENLARGED LIVER




OVARIAN CYST - SMOOTH SOFT MASS




ENLARGED KIDNEY




CARCINOMA ON MASS IN LEFT ILIAC FOSSA



CARCINOMA WITH LIVER METASTASIS



LYMPHOMA



AORTIC ANEURYSM

